

## Manual

# chlori::lyser V1.1

October 2017 Release





# Table of Contents

<b>1</b>	<b>General</b>	<b>5</b>
<b>2</b>	<b>Safety Guidelines</b>	<b>6</b>
2.1	Declaration of Conformity	6
2.2	Special Hazard Warning	6
<b>3</b>	<b>Technical Description</b>	<b>7</b>
3.1	Intended Use	7
3.2	Functional Principle	7
3.3	Product	8
3.4	Storage and Transport	10
3.5	Scope of Delivery	10
3.6	Product Updates, Other	10
<b>4</b>	<b>Installation</b>	<b>11</b>
4.1	Environment	11
4.2	Assembling of chlori::lyser	11
4.3	Mounting in Flow Cell Tap Water	13
4.4	Mounting in Flow Cell Waste Water	14
4.5	Mounting with Probe Carrier	15
<b>5</b>	<b>Initial Startup</b>	<b>16</b>
5.1	Controller for Operation	16
5.2	Connection to the Controller	16
5.3	Probe Initialisation	16
5.3.1	Probe Initialisation using con::lyte D-31x	16
5.3.2	Probe Initialisation using con::lyte D-320	17
5.3.3	Probe Initialisation using moni::tool	18
5.4	Probe Parameterisation	19
5.4.1	Probe Parameterisation using con::lyte D-31x	19
5.4.2	Probe Parameterisation using con::lyte D-320	19
5.4.3	Probe Parameterisation using moni::tool	21
<b>6</b>	<b>Calibration</b>	<b>22</b>
6.1	Types of Calibration	22
6.1.1	Slope Calibration	22
6.2	Performing a Calibration	23
6.2.1	Calibration using con::lyte D-31x	23
6.2.2	Calibration using con::lyte D-320	24
6.2.3	Calibration using moni::tool	25
<b>7</b>	<b>Data Management</b>	<b>26</b>
7.1	Data Storage	26
7.2	Data Transfer	26
7.3	Data Visualisation	26
<b>8</b>	<b>Functional Check</b>	<b>27</b>
8.1	Check of System / Monitoring Station	27
8.2	Check of Results	28
8.3	Check of Probe - Sensor Integrity	29

<b>9</b>	<b>Maintenance</b>	<b>30</b>
9.1	Cleaning of Membrane Cap	30
9.2	Replacement of Electrolyte and Membrane Cap	30
<b>10</b>	<b>Troubleshooting</b>	<b>32</b>
10.1	Typical Error Pattern	32
10.2	Error Messages and Status Messages	33
10.3	Device Settings	36
10.3.1	Check of Device Settings using con::lyte D-31x	36
10.3.2	Check of Device Settings using con::lyte D-320	36
10.3.3	Check of Device Settings using moni::tool	37
10.3.4	Check of Device Settings using ::pro	37
10.4	Return Consignment (RMA)	38
<b>11</b>	<b>Accessories</b>	<b>39</b>
11.1	Installation	39
11.1.1	Connection Cable	39
11.1.2	Extension Cable	39
11.1.3	Sensor Mounting	39
11.1.4	Flow Cell Setup Tap Water	40
11.1.5	Flow Cell Setup Waste Water	41
11.1.6	System Panel micro::station / nano::station	41
11.2	Spare Parts	42
11.2.1	Servicset for free Chlorine (E-520-1/2-KIT)	42
11.2.2	Servicset for total Chlorine (E-525-1/2-KIT)	42
11.2.3	Electrolyte for free Chlorine (E-507-1/2-EL)	42
11.2.4	Electrolyte for total Chlorine (E-507-3/4-EL)	42
11.2.5	Membrane Cap for free Chlorine (E-507-1/2-SET)	43
11.2.6	Membrane Cap for total Chlorine (E-507-3/4-SET)	43
<b>12</b>	<b>Technical Specifications</b>	<b>44</b>

# 1 General

This manual contains, firstly, general information (chapter 1) and safety guidelines (chapter 2). The next chapter (chapter 3) provides a technical description of the s::can product itself as well as information regarding transport and storage of the product. In further chapters the installation (chapter 4) and the initial startup (chapter 5) are explained. Furthermore information regarding calibration of the device (chapter 6), data management (chapter 7), how to perform a functional check (chapter 8) and maintenance (chapter 9) can be found in this manual. Information regarding troubleshooting (chapter 10), the available accessories (chapter 11) and the technical specifications (chapter 12) complete the document.


Each term in this document that is marked *italic and underlined*, can be found on the display of your controller or as lettering on your s::can product.

In spite of careful elaboration this manual may contain errors or incompleteness. s::can does not assume liability for errors or loss of data due to such faults in the manual. The original manual is published in English and German by s::can. This original manual serves as the reference in case discrepancies occur in versions of the manual after translation into third languages.

This manual and all information and figures contained therein are copyrighted. All rights (publishing, reproduction, printing, translation, storage) are reserved by s::can Messtechnik GmbH. Each reproduction or utilisation outside the permitted limits of the copyright law is not allowed without previous written consent from s::can Messtechnik GmbH. The reproduction of product names, registered trade names, designation of goods etc. in this manual does not imply that these names can be used freely by everyone; often these are registered trade marks, even if they are not marked as such.

This manual, at the time of its publication (see release date printed on the top of this document), concerns the s::can products listed in chapter 3. Information and technical specifications regarding these items in s::can manuals from earlier release dates are herewith replaced by this manual.

The electronic version (pdf-document) of this manual is available on the Customer Portal (Services for Customer) of the s::can Homepage ([www.s-can.at](http://www.s-can.at)).



## 2 Safety Guidelines

Installation, electrical connection, initial startup, operation and maintenance of any s::can product as well as complete s::can measuring systems must only be performed by qualified personnel. This qualified personnel has to be trained and authorised by the plant operator or by s::can for these activities. The qualified personnel must have read and understood this manual and have to follow the instructions contained in this manual.



For proper initial startup of complete s::can measuring systems, the manuals for the controller and software used for operation (e.g. con::lyte, con::cube, con::nect, moni::tool), the connected probes and sensors as well as the used additional devices (e.g. compressor) have to be consulted.

The operator has to obtain the local operating permits and has to comply with the joint constraints associated with these. Additionally, the local legal requirements have to be observed (e.g. regarding safety of personnel and means of labour, disposal of products and materials, cleaning, environmental constraints). Before putting the measuring device into operation, the operator has to ensure that during mounting and initial startup – in case they are executed by the operator himself – the local legislation and requirements (e.g. regarding electrical connection) are observed.

All s::can products are leaving our factory in immaculate technical and safety conditions. Inappropriate or not intended use of the product, however, can cause danger! The manufacturer is not responsible for damage caused by incorrect or unauthorised use. Any kind of manipulation of the instrument is strictly prohibited - except for the activities described in this document. Conversions and changes to the device must not be made, otherwise all certifications and guarantee / warranty become invalid. For details regarding guarantee and warranty please refer to our general conditions of business.

### 2.1 Declaration of Conformity

This s::can product has been developed, tested and manufactured for electromagnetic compatibility (EMC) and according to applicable European standards, as defined in the declaration of conformity.

CE-marks are applied on the device. The declaration of conformity related to this marking can be requested from s::can or your local s::can sales partner or can be downloaded from the s::can Customer Portal.

### 2.2 Special Hazard Warning



Because the s::can measuring systems are frequently installed in industrial and communal waste water applications, one has to take care during mounting and demounting of the system, as parts of the device can be contaminated with dangerous chemicals or pathogenic germs. All necessary precautions should be taken to prevent endangering of one's health during work with the measuring device.



Some electrolytes contain diluted acids. Do not swallow the electrolyte. Avoid contact of the electrolyte with skin and eyes. Otherwise wash with a lot of water. In case of eye inflammation, contact a doctor.

## 3 Technical Description

### 3.1 Intended Use

The chlori::lyser is an electrochemical based sensor designed for the continuous monitoring of free or total chlorine in water. For this purpose two different sensor types are available. The value is expressed in mg/l. All sensor types provide the measured temperature as an additional parameter.



The sensors are not suitable for checking the absence of chlorine.

These sensors were developed for use in drinking water, pool water and different types of water treatment. The usage in waste water is possible, but has to be evaluated in the specific application. These sensors have only a low dependence on fluctuations of the measuring flow. Nevertheless a constant flow of the measuring medium is recommended.

In all types of applications, the respective acceptable limits, which are provided in the technical specifications in the respective s::can manuals, have to be observed. All applications falling outside of these limits, and which are not authorised by s::can Messtechnik GmbH in written form, do not fall under the manufacturer's liability.

The device must only be used for the purpose described in this manual. Use in applications not described in this manual, or modification of the device without written agreement from s::can, is not allowed. s::can is not liable for claims following from such unauthorised use. In such a case, the risks are the sole responsibility of the operator.


### 3.2 Functional Principle

The chlori::lyser is a membrane covered amperometric 3 electrode sensor. The hydrophilic membrane, which is not ion specific but permeable to ionic species in general, is stretched over a gold cathode (working electrode). A reference electrode (silver / silver chloride) and a counter electrode (stainless steel), that is special placed on the outside measuring cell, complete the electric circuit.

Free chlorine is defined as the concentration of residual chlorine in water present as dissolved gas ( $\text{Cl}_2$ ), hypochlorous acid (HOCl), and / or hypochlorite ion ( $\text{OCl}^-$ ). These three forms of free chlorine are existent in a pH-dependent equilibrium (see orange line in the figure on the right). The sensor can be used in applications where chlorine gas  $\text{Cl}_2(\text{g})$ , sodium hypochlorite  $\text{NaClO}$ , calcium hypochlorite  $\text{Ca}(\text{OCl})_2$  or electrically generated chlorine are used as disinfectants.

Total chlorine is the sum of free chlorine and combined chlorine (e.g. chloramines).

Due to an internal pH correction, the chlori::lyser reduces the pH dependence of the measurements. As a result, variations of the pH value only have a small influence on the measured value (see blue line in the figure on the right).



### 3.3 Product

The following device variants of the chlori::lyser are available. Regarding detailed information of the device variants please refer to the technical specifications located at the end of this manual.

Type	Specification	Range
E-520-1-000 [E-507-1-000 <sup>1)</sup> ]	chlori::lyser for free chlorine (FCL) with plug connection	0 - 2 mg/l
E-520-2-000 [E-507-2-000 <sup>1)</sup> ]	chlori::lyser for free chlorine (FCL) with plug connection	0 - 20 mg/l
E-525-1-000 [E-507-3-000 <sup>1)</sup> ]	chlori::lyser for total chlorine (TCL) with plug connection	0 - 2 mg/l
E-525-2-000 [E-507-4-000 <sup>1)</sup> ]	chlori::lyser for total chlorine (TCL) with plug connection	0 - 20 mg/l
E-520-1/2-KIT	Electrolyte and membrane cap for free chlorine (spare part for E-520-x-000)	
E-525-1/2-KIT	Electrolyte and membrane cap for total chlorine (spare part for E-525-x-000)	
E-507-1/2-EL	Electrolyte for free chlorine (spare part for E-507-1-xxx and E-507-2-xxx)	
E-507-3/4-EL	Electrolyte for total chlorine (spare part for E-507-3-xxx and E-507-4-xxx)	
E-507-1/2-SET	Membrane cap for free chlorine (spare part for E-507-1-xxx and E-507-2-xxx)	
E-507-3/4-SET	Membrane cap for total chlorine (spare part for E-507-3-xxx and E-507-4-xxx)	

<sup>1)</sup> Previous version of chlori::lyser (has been delivered from s::can until April 2017)

The device is typified by a type label, as shown on the right, that contains the following information:

- Manufacturer's name and country of origin
- Several certification marks
- Device name
- Measuring range
- Bar code
- Device serial number (S/N)
- Information on power supply
- Acceptable temperature limits
- Environment rating (IP)
- Acceptable pressure limits
- Item number (Type)
- QR code to s::can Support



scan Messtechnik GmbH  
Brigittagasse 22-24,A-1200

**chlori::lyser**

Total Chlorine  
0 - 2 mg/l



S/N: 17281605


Made in  
AUSTRIA

9 - 30 VDC  
0 - 45 °C, IP67  
max. 3.0 bar

Type:  
E-525-1-000




- 1 Connector for sensor cable
- 2 Sensor housing
- 3 Counter electrode
- 4 Membrane cap
- 5 Rubber band (to protect vent hole)
- 6 Membrane
- 7 Reference electrode
- 8 Electrode finger
- 9 Goldelectrode



Dimensions of chlори::lyser in mm

### 3.4 Storage and Transport

The temperature limits for device storage and transport, which are described in the section technical specifications, have to be observed at all times. The device shall not be exposed to strong impacts, mechanical loads or vibrations. The device should be kept free of corrosive or organic solvent vapours, nuclear radiation as well as strong electromagnetic radiation.

For short term storage (up to 24 hours) the sensor can stay in the flow cell with water or can be covered with the protective cap filled with water to prevent the sensor from drying out.

For long term storage the membrane cap has to be unscrewed. Then the membrane cap, the spacer and electrode must be rinsed with clean water. Finally the dry membrane is screwed onto the sensor loosely, to protect the electrode finger. The membrane itself should not touch the electrode finger. For recommissioning after long term storage please refer to section 9.2.

Damage to the sensor caused by wrong storage will not be covered by warranty.

Membrane caps that have been in operation for more than one day cannot be used again after storage.

Transport should be done in a packaging that protects the device (original packaging or protective covering if possible).



This product is marked with the WEEE symbol to comply with the European Union's Waste Electrical & Electronic Equipment (WEEE) Directive 2012/19/EC. The symbol indicates that this product should not be treated as household waste. It must be disposed and recycled as electronic waste. Please assist to keep our environment clean.

### 3.5 Scope of Delivery

Immediately upon receipt, please check the received consignment for completeness on the basis of the delivery note and check for any possible damage incurred during shipping. Please inform the delivering dispatcher and s::can immediately in case of any damages in transit.

The following parts should be included in the delivery:

- s::can chlori::lyser (part-no. E-520-x-000 or E-525-x-000), electrolyte and membrane cap are packed separately
- Connection cable (part-no. C-1-010-sensor),
- s::can manual chlori::lyser (part-no. S-295-m)

The following parts could be included in the delivery if ordered as an option:

- Extension cable (part-no. C-210-sensor, C-220-sensor or C-230-sensor)
- Flow cell setup tap water for single sensor (part-no. F-45-sensor)
- Flow cell setup tap water for four sensors (part-no. F-45-four)
- Flow cell setup tap water for i::scan and three sensors (part-no. F-46-four-iscan)
- Carrier for single sensor (part-no. F-12-sensor)
- Automatic flow restrictor (part-no. F-45-flow-1)
- Maintenance set for free chlorine (part-no. E-520-1/2-KIT)
- Maintenance set for total chlorine (part-no. E-525-1/2-KIT)

In case of incompleteness please contact your s::can sales partner immediately!

### 3.6 Product Updates, Other

The manufacturer reserves the rights to implement, without prior notice, technical developments and modifications in the light of continuous product care.

## 4 Installation

### 4.1 Environment

The correct installation of measuring instruments is an important prerequisite for satisfactory operation. Therefore the following checklist for the installation can be used to ensure that all sources for potential operational problems can be ruled out to the greatest possible extent during the installation, allowing the monitoring system to operate properly.

- Favourable flow conditions (little turbulence, acceptable flow rate, pressure, etc.)
- Unadulterated, representative measuring medium
- Measuring medium is in equilibrium state (no gas release, no precipitation, etc.)
- No external interferences (no electric and electro-magnetic interferences by leakage current, earth fault of pumps, electric motors, electric power lines, etc.)
- Easy accessibility (mounting, sampling, functional check, demounting)
- Availability of sufficient space (probe / sensor, installation fitting, controller, etc.)
- Adherence to limit values (see technical specifications located at the end of this manual)
  
- Power supply for controller (operational reliability, voltage, power, peak free)
- Best possible weather and splash water proof conditions
- Shortest possible distances between system components (probe / sensor – controller – compressed-air supply – energy supply)
- Correct dimensioning, mounting and protection of all cables and lines (non-buckling, no risk of stumbling, no damage etc.)

### 4.2 Assembling of chlori::lyser

The chlori::lyser is not ready for use after shipment. This sections will guide you through the steps needed to prepare chlori::lyser for measurement. Please mind the following important notes when handling with the chlori::lyser:

- Never touch the surface of the membrane, the electrode finger or the gold electrode at the tip of it with your fingers.
- Do not shake the electrolyte bottle, store it always upside-down and fill the membrane cap slowly to avoid air bubbles within the electrolyte. Air bubbles between gold electrode and membrane will falsify your readings.
- Do not cover the vent hole with your fingers, when the membrane cap is screwed on or off.



- Please note the correct handling of electrolyte bottle during filling, to avoid air bubbles entering the electrolyte

- 1 Press bottle when upside down for filling.
- 2 Turn your hand to brings the bottle upright.
- 3 Only now open hand to let air flow into the bottle.



**1** Place the sensor package and a clean plastic sheet onto a flat table. Carefully put the provided parts (sensor with membrane cap and protective cap, electrolyte and, if included, spacer) onto the clean plastic sheet.



**2** Place the bottle of electrolyte upside down. Ideally some days before.



**3** Unscrew the membrane cap from the sensor body.



**4** Remove the protective cap carefully. Don't touch the membrane.



**5** Place the membrane cap onto the clean plastic with the opening upwards.



**6** Take the bottle with electrolyte and open it and keep it always upside down. Let the electrolyte flow onto the plastic sheet. Don't stop the flow.



**7** For previous sensor version E-507-1/2 only the following step is needed:  
Put a drop of the electrolyte on the plastic sheet, place the spacer on it and fill the spacer with electrolyte.



**8** Now move the bottle over the membrane cap without stopping the flow. Let the electrolyte flow slowly along the edge into the membrane cap up to the top. This procedure will avoid air bubbles within the electrolyte.



**9** For previous sensor version E-507-1/2 only the following step is needed:  
Hold the sensor body upright and push the electrode tip carefully into the filled spacer to take it up.



**10** Hold the sensor upright and immerse the electrode tip slowly into the filled membrane cap.



**11** Screw the membrane cap onto the sensor body. Electrolyte will escape through the vent hole while screwing. Be careful not to block this vent with your fingers.



**12** Spill the excessive electrolyte from the sensor body with clean tap water or distilled water.



Once the sensor has been assembled, it should be supplied with power as soon as possible and ensure that the sensor is always wetted with disinfected water.

### 4.3 Mounting in Flow Cell Tap Water

This section explains how the chlori::lyser can be installed in a flow cell using either the single sensor flow cell setup (part-no. F-45-sensor) or the four sensor flow cell setup (part-no. F-45-four or part-no. F-46-four-iscan). Regarding the dimensions of these flow cells please refer to section 11.1.4.


After the chlori::lyser is prepared as described in section 4.2 the installation in this flow cell is performed by the following steps:

 Please note the chlori::lyser has to be placed into an opening located after the flow restrictor (i.e. position 3 or 4 in case of F-46-four-iscan).

- Pull out the metal bracket [1] from the flow cell [2] that fixes the plug [3]. A flat screw driver [4] can be used to do this, if needed.



- Remove the plug [3] from the flow cell [2]. To remove the plug insert a flat screw driver [4] into the small hole on the side of the flow cell and move the plug out by moving the screw driver downwards.



- Insert sensor [5] in the opening of the flow cell [2] and push sensor down carefully until O-ring snaps into the correct sensor position.
- Push the metal bracket [1] back into the flow cell [2] to secure the sensor [5] in place. The metal bracket can only be inserted if sensor is in the correct position.
- Ensure that all other openings of the flow cell [2] are covered with plugs [3] before putting the monitoring station into operation.
- To demount the sensor [5] use a flat screw driver [4] to remove the metal bracket [1] first and pull the sensor out.

The flow cell [2] is mounted onto the panel [7] of the micro- / nano::station with two fixing holders [6]. The position of the flow cell is secured by a metal bracket [1].



#### 4.4 Mounting in Flow Cell Waste Water


This section explains how the chlori::lyser can be installed in the flow cell for waste water (part-no. F-48-sensor). Regarding the dimensions of this flow cell please refer to section 11.1.5.



Please note that all specifications regarding flow conditions are kept in mind to receive good measurements

After the chlori::lyser is prepared as described in section 4.2 the installation in this flow cell is performed by the following steps:

- Remove the protective cap from the membrane.
- Pull out the metal bracket with a screw driver as explained in section 4.3.
- Insert sensor [1] in the opening of the flow cell setup [2] and push sensor down carefully until O-ring snaps into the correct sensor position.
- Push the metal bracket [3] into the two holes on the side of the flow cell setup to secure the sensor in place. The metal bracket can only be inserted if sensor is in the correct position.
- Connect pipes / tubes for medium supply to the inlet and to the outlet of the flow cell setup (please refer to section 11.1.5 for dimension of fittings).
- To demount the sensor use a flat screw driver to remove the metal bracket first and pull the sensor out.



## 4.5 Mounting with Probe Carrier

This section explains how the chlori::lyser can be installed in the probe carrier (part-no. F-12-sensor). Regarding the dimensions of this installation accessories please refer to section 11.1.3.




Please note that all specifications regarding flow conditions are kept in mind to receive good measurements.

After the chlori::lyser is prepared as described in section 4.2 the installation in this flow cell is performed by the following steps:

- Remove retaining clip [2] from the probe carrier [1].



- Put extension pipe OD 50 mm or 1 1/2 inch [3] - to be provided by the customer - into the probe carrier [1].
- Drill two holes into the correctly positioned extension pipe [3]. Use the two existing holes [6] for the retaining clip [2] in the probe carrier [1] as guiding help.
- Snap the retaining clip [2] into both holes [6]. Doing this the probe carrier [1] will be fixed onto the extension pipe [3].



Depending on the extension pipe's OD you can use one of the two O-rings [4], included in delivery, to stabilize the position of the pipe. O-ring 50 x 2.5 mm can be used for 50 mm and O-ring 50 x 3.5 mm can be used for 1 1/2 inch.

- Lead the probe cable and the air hose for automatic probe cleaning through the probe carrier (see left figure below).
- Push the sensor into the probe carrier (see middle figure below).
- Tight the screw [5] on the probe carrier using a screw driver until the sensor is firmly fixed (see right figure below).



## 5 Initial Startup

Once the assembling, mounting and installation of the chlori::lyser have been completed and checked (see chapter 4) the initial startup of the s::can monitoring system will require the following actions, in the order presented below:

- Connect the chlori::lyser to the controller used for operation (see section 5.1 and 5.2).
- Establish power supply to the controller (please refer to the manual of the controller) and wait until the operation software has started up.
- Perform probe initialisation of the chlori::lyser. Refer to section 5.3.1 in case of using a con::lyte D-31x, refer to section 5.3.2 in case of using con::lyte D-320 and refer to section 5.3.3 in case of using moni::tool.
- Perform parameterisation of the chlori::lyser. Refer to section 5.4.1 in case of using a con::lyte D-31x, refer to section 5.4.2 in case of using con::lyte D-320 and refer to section 5.4.3 in case of of using moni::tool.
- Configure the measuring interval; additional information can be found in the respective manual of the controller.
- In case required, configure the digital, analogue and fieldbus outputs of the controller.
- Check the readings obtained for plausibility after sufficient running-in time (see section 12 regarding running-in time).
- If necessary calibrate the readings of the chlori::lyser in stable water quality (see chapter 6).

### 5.1 Controller for Operation

For proper operation of the chlori::lyser you will need one of the following controller and operating software respectively.

Controller	Type	Software
con::lyte	D-318, D-319	V5.01 or higher
con::lyte	D-320	V6 or higher
con::cube	D-315	moni::tool V1.3 or higher



s::can recommends to use the most actual version of the operating software on the controller. For service operation with ana::pro please refer to section 10.3.4.

### 5.2 Connection to the Controller

The chlori::lyser will be delivered with a plug connection on the sensor itself only, the connection cable C-1-010 has to be used to connect the chlori::lyser to a compatible socket provided on the controller. Ensure that the sensor plug and the connector are dry and clean. Otherwise communication errors and / or device damage might occur.

In case the controller does not supply enough sockets the distribution box for sensors C-41-hub can be used.

### 5.3 Probe Initialisation

For operating one or several probes using one operation terminal it is necessary to allocate an individual address to every probe. This can be done during probe initialisation at which the probe has to be recongnized first by the ontroller for operation and then a modification of the actual (preset) probe address might be performed. The corresponding address will be stored on the respective probe. For s::can probes and sensors of the same type the same address is preset ex factory.

#### 5.3.1 Probe Initialisation using con::lyte D-31x

The con::lyte should not be powered down or switched off during the initialisation process. In case of rebooting of the con::lyte during the initialisation process (e.g. caused by loss of power supply) the complete procedure of probe initialisation has to be repeated.



- Establish the power supply to the con::lyte and select entry Settings / Parameter settings / Install Probes in the main menu.
- Connect the chlori::lyser to the con::lyte (see section 5.2).
- Push the button Enter, which starts the automatic search procedure for the connected probe. Once the probe is found, address 1 will be allocated. This procedure can last several seconds (see figures below).
- The successful completion of the initialisation will be displayed over a user message. If this message is displayed the initialisation procedure can be finished by pushing the button Esc.

```

Install probe 1
Connect only
probe 1
Continue with ENTER
Stop with ESC
    
```

```

Install probe 1
Searching for probe
    
```

```

Install probe 1
Probe search finished
chlori::lyser found
Continue with ENTER
Stop with ESC
    
```

A user message will also be displayed when no probe is detected. In this case please check the following before repeating the procedure for probe initialisation:

- Is only one probe connected to the con::lyte?
- Is the probe connected properly?
- Are all wires of the con::lyte socket in the terminal compartment tight?

```

Install probe 1
Probe search finished
No probe found
Continue with ENTER
Stop with ESC
    
```

### 5.3.2 Probe Initialisation using con::lyte D-320

At the initial start-up the con::lyte D-320 provides an automatic probe and sensor initialisation procedure (see screen on the right). After connecting all probes and sensors to the appropriate plugs of the con::lyte (see section 5.2) and pushing the OK button, the probe and sensor initialisation starts.

```

Add s::can sensor...
Please connect all
sensors and press
OK to continue...
    
```

If chlori::lyser will be initialized at a later date, the following steps are needed:

- Switch to Status display by using the Left- or Right button.
- Push Function button, select menu Manage sensors... and confirm with OK.
- Select menu Add sensor... and confirm with OK.
- Connect chlori::lyser to the D-320 (see section 5.2).
- Select menu Add s::can sensor... and confirm with OK.

```

Add new Sensor
Add 0/4-20mA...
Add digital in...
Add s::can sensor...
    
```

As soon as the entry is confirmed by pushing the OK button, the con::lyte will automatically search the Modbus port for a new sensor and will add the new sensor to the sensor list.

```

Add s::can Sensor...
Searching 17/20
F: chlori::lyser/0/9
A: chlori::lyser/0/9
    
```





After adding a new probe or sensor, the parameters will be displayed in the parameter screen. Furthermore single parameter can be added manually (see section 5.4.2 and menu Add parameters...). In case the installation failed, the message Error adding! will be displayed.

```

Add s::can Sensor...
Done. Press OK...
Added sensors: 1
Replaced sensors: 0
    
```

### 5.3.3 Probe Initialisation using moni::tool

- Click the Service tab of the moni::tool screen and logon as Administrator.
- Click on an empty sensor icon (Add new Sensor) to initiate the initialisation process.
- An automatic search procedure will start, searching for the connected sensor.
- When the automatic search procedure is finished, moni::tool will display all connected probes and sensors. Those sensors that are connected for the first time will have the Status Found new sensor (also listed as New Sensors below).
- If needed Sensor name can be modified now, which can be any descriptive name you desire or select one of the previous names listed below this entry field.
- To install the new sensor click on the blue + sign on the right side of the sensor or push the button Install All.
- moni::tool will install the sensor and switch to the Service tab showing the new sensor in the system overview. Now pushing the button Leave Service Mode located on the upper left side to start the measuring process.
- When pushing the button Advanced Search (see 2<sup>nd</sup> and 3<sup>rd</sup> figure from the top), the method you connected the sensor (Connection methode), the used COM-Port and the Address can be defined. A click on button Start search will start search procedure within the defined range.
- After the Advanced Search is finished, the Install Sensor screen is displayed. Depending on the used Service Level (Basic, Advanced or Expert) only the basic or all information received from the sensor are listed up.
- To install the new sensor push the button Save in the upper part of the window.



## 5.4 Probe Parameterisation

The following table is an overview of the parameters that can be measured with the chlori::lyser:

Sensor / Part-no.	Previous Version	Parameter	Parameter-index	Name [Unit]	Decimal places
chlori::lyser E-520-1	E-507-1	Free chlorine	0	FCL [mg/L]	3
		Temperature	1	Temp. [°C]	1
chlori::lyser E-520-2	E-507-2	Free chlorine	0	FCL [mg/L]	2
		Temperature	1	Temp. [°C]	1
chlori::lyser E-525-1	E-507-3	Total chlorine	0	TCL [mg/L]	3
		Temperature	1	Temp. [°C]	1
chlori::lyser E-525-2	E-507-4	Total chlorine	0	TCL [mg/L]	2
		Temperature	1	Temp. [°C]	1

### 5.4.1 Probe Parameterisation using con::lyte D-31x

After successful probe initialisation (see section 5.3.1) the measuring parameters of the chlori::lyser will be displayed on the display of the con::lyte automatically. If needed the measuring parameters can be configured individually using the menu item *Settings / Parameter settings / Parameter n* (*Settings / Parameterconfig / Parameter n* with older versions).

The name of the *Probe* or sensor used as a source of the parameter is displayed in the upper line (e.g. chlori::lyser). If several probes or sensors are installed the instrument from which a parameter needs to be displayed can be selected here. Under the entry Probe the *Address* that has been allocated to that probe is displayed as an additional information. The *Index* specifies the place of the corresponding parameter onto the allocated probe. The *Unit* of the selected parameter is displayed in the line below. The item *Decimal places* enables settings of the number of displayed decimal places (between 0 and 4). With the default setting *auto* the number of decimal places will be automatically set by the sensor.

Parameter 1	
Probe:	chlori::lyser
Address:	1
Index:	0
Unit:	mg/l
Decimal places:	auto

### 5.4.2 Probe Parameterisation using con::lyte D-320

After successful probe initialisation (see section 5.3.2) the needed measuring parameters of the chlori::lyser have to be added to the parameter display. This is performed by the following steps:

- Switch to status display with *Left-* or *Right* button.
- Push *Function* button, select menu *Manage sensors...* and confirm with *OK*.
- Select *chlori::lyser/0/x* and confirm with *OK*.
- Select menu *Add parameters...* and confirm with *OK*.
- Select needed parameter and confirm with *OK*.

Add para.	
▶ Add	FCL
Add	Temp.

The selected parameter will be displayed now on the next free position of the parameter display. The default display configuration is used. Changing the display format is performed by the following steps:

- Select the parameter in der parameter display with Up- or Down button.
- Push Function button, select menu Display settings... and confirm with OK.
- Select chlori::lyser/0/x and confirm with OK.
- Select menu Add parameters... and confirm with OK.
- Select needed parameter and confirm with OK.

In the displayed parameter configuration the following settings can be modified.

- Name                Displays the actual name of the paramter.
- Unit                 Displays the actual unit of the paramter.

P2 / Temp	
Name:	Temp
Unit:	°C
Disp.Format:	1
Load Defaults	

To change the name or unit of the parameter, select the entry with Up- and Down buttons and by pushing the OK button the name can be changed with Up-, Down-, Left- and Right buttons. Pushing the OK button confirms the new name. Please note that change of parameter name or unit will not change the parameter configuration itself (e.g. if you change the parameter name NO<sub>3</sub>-N to NO<sub>3</sub> the reading will still be NO<sub>3</sub>-N).

- Disp.Format    Within this line the number of displayed decimal places (between 0 and 5) can be set. Please note that in case of too many digits high values can not be displayed and the parameter reading will switch to plus signs (++.+++++).
- Load Defaults        Confirming this entry by pushing the Ok button will restore the default display settings from the sensor.





All modifications performed by the operator within these settings menu will be documented in the logfile of the con::lyte (see manual con::lyte D-320).

### 5.4.3 Probe Parameterisation using moni::tool

After successful probe initialisation (see section 5.3.3) all parameters available on the probe will be installed and automatically displayed on the *Value* screen of moni::tool. If not all new parameters are displayed, please check maximal number of parameters of your moni::tool license. If you want to configure the measuring parameters individually this can be done using the menu item *Service / Terminal / Parameter*.

After selecting that menu item a list of all installed parameters is displayed. After selecting one or several parameters by clicking on them the following activities can be performed:

- Moving the selected parameter to a higher position in the *Value* display by pushing the entry *Up*.
- Moving the selected parameter to a lower position in the *Value* display by pushing the entry *Down*.
- Deleting the selected parameter from *Value* display by pushing the entry *Remove Parameter*.
- A new parameter can be added by pushing the entry *Add Parameter*.
- Click on the blue wheel (*Edit*) on the right hand side of the parameter will display the actual parameter settings. Depending on the actual *Service Level* different settings are displayed and can be edited (*Parametername*, *Unit* and *Resolution*).
- Click on the blue check mark (*::tool*) or on the blue sign (*Alarm*) on the right hand side of the parameter to check or modify the settings for vali::tool and parameter alarm. Please refer to manual moni::tool for further information.



## 6 Calibration

The chlori:lyser is delivered with membrane cap and electrolyte separately. Therefore the electrode slope needs to be calibrated after initial startup (see chapter 5). Subsequently a calibration is needed whenever maintenainces activities (changing the electrolyte or the membrane cap) have been performed.

- Before performing any kind of calibration ensure appropriate conditioning time (at least 2 hour after initial operation).
- Before performing any kind of calibration the correct function of the sensor should be ensured (sensor is clean and properly assembled - see section 8 also).
- For highest accuracy the same environment conditions (temperature, flow velocity, pH) have to be ensured during the calibration as for the normal operation. Therefore the calibration should always be performed with the installed sensor directly in the medium and only in exeptional cases outside the flow cell.
- When calibration is performed outside the flow cell in a small beaker, ensure that the sensor is not in direct contact with the wall or bottom and the outer steel ring of the sensor is submersed in the medium.
- Take a sample from the medium at the same time when pushing the sample button in the calibration menu (actual measurement will be stored).
- Perform laboratory analysis for actual chlorine concentration of the sample as fast as possible.
- s::can recommends to use DPD methodes for free chlorine or total chlorine with liquid reagents as reference method.
- For highest accuracy usage of a photometer for measuring chlorine concentration of the sample is recommended. Perform zero-point calibration of the photometer before usage.
- The calibration will not be executed and used till the corresponding menu item (e.g. *Calibrate!*) is confirmed.
- When performing a parameter calibration the result will be checked for plausibility. In case of faulty calibration an error message will be displayed to the operator. Please refer to section 10.2 regarding possible error messages and notes for removal.

### 6.1 Types of Calibration

For calibration of the free (FCL) or total (TCL) chlorine parameter a slope calibration (SPAN) can be performed. This local calibration adapts the global calibration to the actual monitored medium.


The zero point is precalibrated in the factory and does not need to be recalibrated in the field.

The temperature value of the internal compensation cannot be calibrated.

#### 6.1.1 Slope Calibration

- The chlori:lyser is equipped with a global calibration (reference slope) ex factory. You can switch back to this factory setting at any time.
- The local calibration shall either be performed directly in the medium or in a calibration standard.
- When using a calibration standard the chlorine value of the used standard shall be closed to the measuring range of the application.
- For slope calibration only one sample is needed. This sample will be stored onto the sensor.
- The result of the slope calibration will be stored directly onto the sensor and will be used until a new slope calibration is performed successfully or you switch back to the default slope.
- The slope calibration will be accepted if the global calibration is changed not more than 70% of the original value (*Reference Slope*).

Select calibration type



## 6.2 Performing a Calibration

### 6.2.1 Calibration using con::lyte D-31x

The Calibration entry in the con::lyte main menu leads you into the menu that enables the calibration of the chlori::lyser. When Calibration is selected a password must be entered (password = 1) before the calibration can be started. The next step is selection of the parameter to be calibrated (e.g. FCL) in the selection field Parameter Calib.

Parameter Calib.	
Local cal.:	FCL
Local cal.:	Temp

Now the menu for local calibration will appear as displayed on the right hand side.

As long as chlori::lyser is working with factory calibration (default) the entry Calib. shows global and no Type can be selected.

Local cal.:		FCL
Calib.:	global	
Type:	None	
Set reference slope		
Calibrate!		

To perform a local calibration the entry Calib. has to be changed to local. Then below the entry Type will show Span as possible type of calibration.

The display shows the reading for chlorine actually measured (Value) as well as the chlorine concentration stored on the sensor for calibration (Sample 1). As long as no sample is stored on the sensor the display will show dashes. When confirming the entry Sample 1 by pushing Enter the raw signal of the actually measured (displayed) chlorine concentration will be stored as new sample on the sensor. On the entry Lab the results corresponding (real chlorine concentration) to the readings stored under Sample 1 can be entered here.

Local cal.:		FCL
Calib.:	local	
Type:	Span	
Value:	1.27	
Sample 1:	--.--	
Lab 1:	--.--	
Set reference slope		
Calibrate!		

When the entry Calibrate! is confirmed by pushing Enter, a calibration is performed. Successful calibration is shown in a user message (Please wait and then PLocal calib. saved). If the calibration was not successful (user message Local calib. Error!) the calibration used up to now will be used further on.

When selecting the entry Set reference slope after performing a local calibration, the new slope will be stored as a reference. This reference can be used to compare further slope calibrations and to document ageing of the membrane / electrolyte. The procedure is confirmed by the message Reference slope was set successfully. Continue with ENTER.



The possibility to store a new reference slope is only supported by this s::can operation controller. Therefore it is disadvised to use of this possibility for internal check of electrode ageing and keep the slope set ex factory instead.

For temperature reading no local calibration is possible.

## 6.2.2 Calibration using con::lyte D-320

This operating controller provides, beside normal calibration procedure (see further down), the possibility for a quick calibration call directly from the parameter view. This is performed by following steps:

- Select the parameter in der parameter display with Up- or Down button.
- Push OK button, which directly displays the calibration screen.
- Select Sample 1 and confirm with OK to store the raw signal of the actual reading.
- Take a water sample to analyse real chlorine concentration.
- Enter the result from laboratory analyse into the field Lab 1.
- Select entry Perform Calibration and confirm with OK.
- Leave the calibration screen with Back button.

< V	P1/2	FCL	>
▶	1.02	FCL mg/l	
	8.7	Temp °C	

P1/FCL	
Lab 1:	1.62
Sample 1:	78.11
Perform Calibration	

The advanced local calibration provides extensive possibilities for calibration of measurement parameter. After selecting the parameter in the parameter display, pushing the Function button, selecting the menu Calibrate expert... and pushing the OK button, the calibration screen is displayed.

- Type Two different types of calibration are available: Local or Global. By default Local is selected. This is the normal calibration performed by the operator. As soon as Global is selected an confirmed with OK a reset of the sensor to factory calibration is performed and the actual reading (Value), the raw signal (Private) and the default slope (Span) will be displayed.
- Mode The only mode available is Span which is preset.
- Perform Calibration Confirming this entry by pushing the Ok button will execute the local calibration, using the Lab and Sample values displayed on the calibration screen.
- Value Displays the measured value of the probe or sensor like on the parameter screen also (i.e. using the actual calibration). The value will be updated permanently.
- Private Displays the according raw value (measured power signal of the sensor) of the displayed reading. The value will be updated permanently.
- Lab 1 Within this line the correct value for the measured Sample 1 has to be entered. The entered Lab value can be either the laboratory result of the sample taken or the concentration of the standard solution, which is used for calibration. The unit of the lab value has to be in accordance with the measuring parameter.  
  
An entered Lab value can be deleted by selecting it and pushing the Function button so that it will not be used in the calibration.
- Sample 1 When confirming this entry by pushing the Ok button, a measurement will be performed and stored as sample 1 for the local calibration. The sample for the laboratory should be taken at the same time. The displayed and stored value, which will be used for the calibration might be a raw value (e.g. mV value) and therefore might also be negative. Existing readings (Sample 1 or Sample 2) are overwritten whenever a new measurement is triggered by pushing the Ok button. If no sample measurement was performed or if the measurement was invalid, the message Measure! will be displayed instead of a numerical value.
- Span Displays the used slope of the actual calibration. It is not possible to edit this value.

P1/FCL	
Type:	Global
Value:	1.02
Private:	9.74
Span:	10.00

P1/FCL	
Typ:	Lokal
Mode:	Span
Perform Calibration	
Value:	1.02
Private:	9.74
Lab 1:	25.3
Sample 1:	-65.8
Span:	6.00

### 6.2.3 Calibration using moni::tool

- Click the Service tab of the moni::tool screen and logon as Administrator.
- Click the icon of the probe / sensor you want to calibrate in the displayed system overview.
- Click the icon Calibrate sensor in the next screen.

Now the screen shows a list of all parameters being measured by this probe / sensor. Clicking on the blue triangles will open more information about actual used calibration for this parameter.

Service > chlo 14450003 > Calibration			
Parameter name	Last calibration	Calibrate	History
Free Chlorine	Administrator [ Span ] Coefficient 0 - Offset: 0.0 Coefficient 1 - Slope: 2.6869		

Furthermore a click on the History icon rightmost opens a logbook showing all up to now with this con::cube performed calibration procedures.

Service > chlo 14450003 > Calibration > Calibrate Free Chlorine

Show sample graph | Perform calibration 5

Open the calibration screen by clicking on the Calibrate icon on the right hand side of the parameter you want to calibrate.

**Calibrate Free Chlorine [mg/l]**

Span 1

Current value: 0.900 mg/l

2

**STATUS**

Parameter: OK

**SAMPLES**

#	Sample	Measured	Laboratory	Edit
1		0.335 mg/l	0.900 mg/l	

3 4

**CURRENT COEFFICIENTS**

Name	Value
Coefficient 0 - Offset	0.0
Coefficient 1 - Slope	2.6869

11:42 11:43 11:44 11:45

This button displays the actual used calibration (Global or Span). Push this button to select the type of calibration you want to perform (see section 6.1).

Watch the current readings being displayed numerically and graphically on the calibration screen. Wait until readings are stable.

Push the Sample icon to store the actual reading onto the sensor. Please note that the displayed value is the raw value (mV reading).

Push the Edit icon to enter the result of the laboratory analysis and store it onto the sensor.

Push the button Perform calibration to start the calibration procedure.

After the calibration procedure is finished a user message will inform you, if the local calibration was successful. The new calibration coefficients will be displayed also. In case of an error the reason will be displayed as well as possible solutions and the detailed error code.

#### Calibration

**New calibration was saved.**

Offset: 0.0  
Slope: 2.6869

Status: OK

OK

#### Calibration

**New calibration was saved.**

Offset: 0.0  
Slope: 74.6369

**Error during calibration, see status:**

Sensor maintenance required  
Contact your local s::can sales partner.  
Reading out of measuring range  
Check whether sensor is in the medium. If yes, perform functional check of the instrument and/or recalibrate sensor using samples with higher and/or lower concentrations.  
Status: 0x0000.8000.0000.8001.0000

OK

## 7 Data Management

### 7.1 Data Storage

The following information are stored directly on the sensor:

- Result of slope calibration
- Default slope
- Measured sample used for slope calibration
- Device information (e.g. type, serialnumber, address, please refer to section 10.3)

There is no possibility to store readings onto the sensor itself.

### 7.2 Data Transfer

The measurements are performed on the sensor and the readings are transferred to the controller via the sensor cable using RS 485.

### 7.3 Data Visualisation

For visualisation of the sensor readings one of the following s::can controller can be used:

- con::lyte
- con::cube
- con::nect with PC

## 8 Functional Check

A functional check might be required for one of the following reasons:

- Initial startup
- Routine functional check
- Suspicion of monitoring system malfunction
- Modification of monitoring system (e.g. integration of additional sensor or device)
- Change of measuring location

Depending on the application (water composition), the probes and the sensors connected and the environmental conditions a regular functional check (weekly to monthly) is recommended. The following sections provide an overview of all the actions that have to be performed to check the monitoring system quickly (see section 8.1), to check the plausibility of the collected readings (see section 8.2) and to check the integrity of a single probe or sensor (see section 8.3).

### 8.1 Check of System / Monitoring Station

Check	con::lyte	moni::tool / con::cube	What to do if check failed
Power supply controller	Green LED is on and text is visible on the display?	LED on housing cover is on or at least flashing? moni::tool screen is displayed after touching the screen?	Check power supply of controller. Power off controller for 5 minutes and power on again.
System running (up-to-date)	Displayed system time is current and is updated every second?	Click on system clock at the bottom of the screen shows current time and last measurement. Both are current?	Check for displayed error messages. Check if Service mode is activated or automatic measurement is paused.
System status	No error messages or error symbols are displayed?	LED of con::cube is blue and <i>Status</i> icon of moni::tool is not blinking yellow?	See section 10 for Troubleshooting.
Reason for bad system status	Check logbook entries since last functional check.	Open <i>Status</i> tab and select symbol of affected sensor for more information.	See section 10 for Status- and Errorcodes.

Check	Remark
Function of automatic cleaning	Use function <i>Clean now</i> or wait for next cleaning cycle. Watch for air bubbles when cleaning is activated.
Compressed air supply for automatic cleaning	All tubes and fittings are tight?
Function of compressor and storage tank	Drain condensed water from storage tank of compressor (not necessary for s::can compressor B-32). Check pressure.
Monitoring station (by-pass)	All tubes and fittings are tight and all probes and sensors are supplied with medium? No air bubbles within the tubes?
Installation submersed (in-situ)	Mounting equipment of all devices is ok and all probes and sensors are submersed?
Data transfer	Check if displayed readings on local controller are equal with displayed readings on customer display system.


## 8.2 Check of Results

Check	con::lyte	moni::tool	ana::xxx
Current readings displayed completely	No <u>NaN</u> and no dashes (- - -, - -) or plus sign (++++,++) displayed. Use arrow buttons to scroll through all displayed parameters.	No <u>NaN</u> displayed.	No <u>NaN</u> displayed.
Current parameter status of displayed readings	Check logbook entries since last functional check.	Red background for parameter indicates an error or alarm. Grey background indicates reading is not current.	Activate <u>Show context help</u> if grey background or <u>NaN</u> and move cursor over displayed reading.

Check	Reason	Remark
Up-to-date: Readings actualised on regulary base?	<ul style="list-style-type: none"> <li>- Measuring interval is too long</li> <li>- Automatic measurement has been stopped manually</li> </ul>	Consider measuring interval and smoothing.
Continuity: Check historical data (timeseries) for inter-ruptions or discontinuities	<ul style="list-style-type: none"> <li>- Change of medium</li> <li>- Local calibration</li> <li>- Maintenance of probe / sensor (cleaning, etc.)</li> <li>- Readings out of range</li> <li>- System failure (loss of power, communication error, etc.)</li> </ul>	Only possible if timeseries are availbale.
Plausibility: Timeseries look plausible with daily or seasonal fluctuation	<ul style="list-style-type: none"> <li>- Drift of readings (can be caused by fouling)</li> <li>- Increasing noise (can be caused by flow conditions or fouling)</li> <li>- Fixed readings / no fluctuation</li> </ul>	Check logbook of plant operator if possible.
Measuring range: Readings are within the specified and calibrated measuring range?		Quality of results might be reduced outside the specified range.
Accuracy: Difference between laboratory values and readings of the chlori::lyser	In case of significant difference a slope calibration (span) has to be performed (please refer to section 6.1.2)	To verify the accuracy of the displayed readings only a reliable and validated comparison method has to be used.

### 8.3 Check of Probe - Sensor Integrity

When there is any doubt regarding the integrity of the sensor, please use the following flowchart to check sensor and installation:



1) Use DPD methods for check of real chlorine concentration

2) Note the technical specifications in section 12

## 9 Maintenance



Please note that during any cleaning or maintenance activity of the chlori:lyser the following important notes have to be obtained:

- Never touch the surface of the membrane, the electrode finger or the gold electrode at the tip of it with your fingers.
- Do not use any acids on the electrode finger, the spacer or the housing.
- Do not power off the sensor for more than 24 hours when it is in operation.
- Do not shake the electrolyte bottle, store it always upside-down and fill the membrane cap slowly to avoid air bubbles within the electrolyte. Air bubbles between gold electrode and membrane will falsify your readings.
- Do not cover the vent hole with your fingers, when the membrane cap is screwed on or off.
- Please note the correct handling of electrolyte bottle during filling, to avoid air bubbles entering the electrolyte (see section 4.2)

### 9.1 Cleaning of Membrane Cap

If the membrane has a coating or is covered with organic material, it can be cleaned with hand warm tap water carefully. Please ensure that the rubber band is positioned correctly to avoid any dilution of the electrolyte. If this cleaning procedure will not improve the measurement, the membrane cap and the electrolyte have to be replaced.

### 9.2 Replacement of Electrolyte and Membrane Cap

Electrolyte and membrane cap have to be replaced on regular interval. The maintenance interval depends on the sensor type and the measured medium. The typical maintenance intervals are:

- Replacement of membran cap once per year
- Replacement of electrolyte for actual version (E-520 and E-525) once per year
- Replacement of electrolyte for previous version (E-507) every 3 - 6 months

In case of troubles outside the scheduled maintenance interval (see section 9) the following step by step procedure is recommended:


- Cleaning of membrane cap (see section 9.1).
- Visual check of the electrode finger (see figures below). If electrode finger looks not ok, please check your application for unexpected aggressive substances.
- Replacement of electrolyte and cleaning of gold electrode (tip of electrode finger), as describes afterwards.
- Replacement of electrolyte and membran cap and cleaning of gold electrode (tip of electrode finger), as described afterwards.



The electrolyte and membrane cap need to be replaced every 12 months or if the local calibration failed (see section 6). The replacement is performed by the following steps:

**1** Clean the sensor as explained in section 9.1.

**2** Lift the rubber band so that the vent hole is not covered when unscrewing the cap.



**3** Unscrew the membrane cap.

**4** Empty the used electrolyte. For previous sensor version E-507-1/2 only take care not to lose the spacer in case only electrolyte will be replaced.





**5** In case the electrolyte has to be changed only, spill the membrane cap with clean tap water or distilled water before refilling.

**6** Clean the electrode finger by rinsing with clean tap water or distilled water without touching it directly.



**7** Shake excessive water off the electrode finger. Do not use tissues to dry the electrode finger or the inner of the membrane cap.

**8** Place the supplied fine polishing paper onto a soft surface (e.g. a paper tissue) and fix it with your fingers. Clean the gold electrode by softly wiping the electrode finger three times over the fine polishing paper. Keep the sensor upright.



**9** Perform refilling with electrolyte and fixing of the membrane cap as explained in section 4.2.

# 10 Troubleshooting

## 10.1 Typical Error Pattern

Error	Reason	Removal
Drift of readings	<ul style="list-style-type: none"> <li>■ Change in the medium</li> <li>■ Fouling of the membrane</li> <li>■ Electrolyte aged</li> <li>■ Membrane aged</li> </ul>	<ul style="list-style-type: none"> <li>■ Check measuring medium for plausibility (reference method)</li> <li>■ Check sensor head for cleanliness</li> <li>■ Replace electrolyte</li> <li>■ Replace membrane cap</li> </ul>
Drift of readings after change of electrolyte or membrane cap or after loss of power supply	<ul style="list-style-type: none"> <li>■ Sensor not fully conditioned</li> <li>■ Medium flow insufficient</li> <li>■ Membrane not clean</li> <li>■ Air bubbles in electrolyte</li> <li>■ Electrolyte aged (e.g. due to wrong storage conditions)</li> </ul>	<ul style="list-style-type: none"> <li>■ Condition sensor after recommissioning in the medium (see section 12 for conditioning time)</li> <li>■ Check water supply, check inlet strainer</li> <li>■ Check membrane cap for cleanliness</li> <li>■ Replace electrolyte</li> <li>■ Use a new bottle of electrolyte and store it correctly</li> </ul>
Periodical deviation of the readings (outliers of readings)	<ul style="list-style-type: none"> <li>■ Periodical fluctuation of pressure</li> <li>■ Attaching and detaching of air bubbles on the membrane</li> <li>■ Periodical fluctuation of flow</li> </ul>	<ul style="list-style-type: none"> <li>■ Ensure pressure conditions are stable (use flow restrictor F-45-flow-1)</li> <li>■ Ensure no air bubbles are within the tubes and flow cells</li> <li>■ Check medium supply and installation</li> </ul>
Unstable readings (scattering of readings)	<ul style="list-style-type: none"> <li>■ Air bubbles on the membrane</li> <li>■ Air bubbles in the electrolyte</li> <li>■ Membrane destroyed</li> </ul>	<ul style="list-style-type: none"> <li>■ Increase flow rate for short time to remove all air bubbles and ensure stable medium flow without bubbles</li> <li>■ Replace electrolyte and do refilling carefully</li> <li>■ Check membrane visually and replace it</li> </ul>
No response to changes in concentration	<ul style="list-style-type: none"> <li>■ No power supply for more that 24 hours</li> <li>■ Longterm measurement in medium without chlorine</li> <li>■ Flow too low</li> <li>■ Electrolyte aged</li> <li>■ Membrane aged</li> </ul>	<ul style="list-style-type: none"> <li>■ Clean membrane cap and replace electrolyte</li> <li>■ Store sensor in chlorine medium and check it again after 20 minutes</li> <li>■ Check water supply, check inlet strainer</li> <li>■ Perform calibration or replace electrolyte</li> <li>■ Perform calibration or replace membrane cap</li> </ul>
Measurement results deviate from laboratory results	<ul style="list-style-type: none"> <li>■ Cross-sensitivity of the reference method (e.g. ammonium)</li> <li>■ Incorrect calibration</li> <li>■ Different pH value during calibration and measurement</li> <li>■ Reference methode</li> <li>■ Too low or too high flow</li> </ul>	<ul style="list-style-type: none"> <li>■ Obey limits of the reference method regarding pH and cross sensitivities</li> <li>■ Perform calibration again</li> <li>■ Calibration should be done on the expected mean of the pH value.</li> <li>■ Be aware that all reference methodes are limited in accuracy (e.g. accuracy of DPD methode for free chlorine up to +/- 0.23 mg/l FCL)</li> <li>■ Check water flow</li> </ul>
Sensor cannot be calibrated	<ul style="list-style-type: none"> <li>■ Start-up time too short</li> <li>■ No electrolyte in the membrane cap</li> <li>■ Protective cap still on the sensor head</li> <li>■ Measurement with reference method performed incorrect</li> <li>■ Sensor with filled membrane cap without power for longer time</li> </ul>	<ul style="list-style-type: none"> <li>■ See section 12 for start-up / conditioning time</li> <li>■ Perform correct sensor assembly according to manual</li> <li>■ Remove protectice cap from the sensor head</li> <li>■ Perform lab measurement according the instruction</li> <li>■ Repeat calibration after 24 hours of operation</li> </ul>

## 10.2 Error Messages and Status Messages

During execution of a measurement the monitoring system (system status), the measuring device itself (device status) and the result (parameter status) will be checked for possible errors and for plausibility. In case of an error (status bit will be set from 0 to 1) a user message will be displayed to the operator.

Depending on the used operation controller these messages will be shown on the display (Menu Logbook & data in case of con::lyte D-31x, function Monitor... in case of con::lyte D-320, Status tab in case of moni::tool and Show Context Help and System-Status in case of ana::pro) and also stored within the result files or logfiles. Additional to the user message (general error reason and recommendations for removal) the detailed status code will be displayed either in binary form (0000, 0001, 0010, 0011, 0100, etc.) or as a hex number (0x0001, 0x0002, 0x0004, 0x0008, 0x0010, etc.).



Up to 16 status bits are used for different errors. If several errors occur at the same time the con::lyte and moni::tool will add up all the status bits. This detailed information might be important if you request s::can support. Below you will find examples how to translate these combined hex codes:

Hex	Bin	Bits
0x8000	1000 0000 0000 0000	b15
0x8001	1000 0000 0000 0001	b0, b15
0x4011	0100 0000 0001 0001	b0, b4, b14


Within moni::tool the complete status code of a simple parameter has the following format: 0xTTTT.SSSS.PPPP.pppp.VVVV.vvvv.

Code	Status Type	Remark
0xTTTT	System status sensor	visible in the second column of all moni::tool parameter result files (e.g. Error 0x0010 or Ok 0x0002)
0xSSSS	Sensor status general	valid for all sensors
0xssss	Sensor status private	valid for respective sensor
0xPPPP	Parameter status general	valid for all parameters
0xpppp	Parameter status private	valid for respective parameter
0xVVVV	vali::tool status general	valid for all clean values of vali::tool software
0xvvvv	vali::tool status private	valid for respective clean values of vali::tool software

Within the moni::tool Status tab of the sensor you will see the system status sensor and the sensor status as clear text and as status code (0xTTTT.SSSS.ssss).

Within the moni::tool Status tab of the parameter you will see the parameter status and in case of activated vali::tool the vali::tool status also (0xPPPP.pppp.VVVV.vvvv).

Within the moni::tool results file of the sensor parameter the status (0xTTTT.SSSS.PPPP.pppp) will be stored in the column beside the measured value and the vali::tool status (0xVVVV.vvvv) in the column beside the cleaned value.



The table below shows all errors regarding the operation terminal (system status) when a chlори::lyser is connected incl. the user message, the reason of the error and notes for troubleshooting. If the error can't be removed although the suggested procedure was executed several times, please contact your s::can sales partner.

System Status Error 0xTTTT	Display con::lyte (D-31x / D-320)	Message moni::tool	Reason	Removal
0x0001 - b0	ES007 / COMM! Probe not detected. Check power-supply and connection cable.	No communication between sensor and terminal.	No communication between sensor and operation terminal. Replacement sensor was not installed correctly.	Check sensor cable and connector. Disconnect and reconnect sensor.
0x0002 - b1	0002	Invalid sensor	Sensor serial number has changed.	Connect the pre- viously installed sensor or perform sensor replacement (moni::tool) or new sensor installation (con::lyte).

The table below shows all errors regarding the used sensor incl. the user message, the reason of the error and notes for trouble shooting. If the error can't be removed although the suggested procedure was executed several times please contact your s::can sales partner.

Sensor Status Error 0xSSSS	Display con::lyte (D-31x / D-320)	Message moni::tool	Reason	Removal
0x0001 - b0	ES100 / 0001 Probe reports an error. Call service! Param.Status error. Status Code: ....	General sensor error	Sensor reports error during internal check. At least one internal sensor check failed.	For details see additional status message below. In case no further messages are shown, note the error code and contact your s::can sales partner.
0x0002 - b1	ES101 / 0002 MISUSE Medium temperature. Take probe out of medium, immediately!	SENSOR MISUSE	Operation outside the specification (e.g. temperature too high). This can damage the device permanently.	Take the sensor out of the medium immediately and check environmental conditions.
0x8000 - b15	ES115 / 8000 Device maintenance required Code 8000 0000	Sensor maintenance required	At least one internal sensor check reports a warning.	Perform function check of the sensor according the manual.

The table below shows all errors regarding the measured parameters incl. the user message, the reason of the error and notes for trouble shooting. If the error can't be removed although the suggested procedure was executed several times please contact your s::can sales partner.

Parameter Status Error 0xPPPP	Display con::lyte (D-31x / D-320)	Message moni::tool	Reason	Removal
0x0001 - b0	EP 100 / 0001 Status error. Code: 0001.0000 Details in following log messages.	General parameter error	At least one internal parameter check failed.	Note additional status message below. If no further message is displayed, note the error code and contact your local s::can sales partner.
0x0002 - b1	EP 100 / 0002 Parameter failure, hardware failure	Parameter error, Hardware error	Electrode signal not ok. An electrode is missing, too old or defect.	Check the electrode or replace the electrode.
0x0004 - b2		Parameter error, configuration error	Parameter error, configuration error	Change the local calibration or swirch back to global calibration.
0x0008 - b3		Parameter error, Wrong medium	Sensor outside of the medium or in incorrect medium.	Check supply of medium and medium itself.
0x0010 - b4	EP 100 / 0010 Parameter failure, calibration failure	Parameter error, Incorrect calibration	Invalid sensor configuration. At least one calibration coefficient is invalid.	Check readings and lab values. Set back to factory settings. Repeat local calibration.
0x0020 - b5	EP 100 / 0020	Parameter not ready	Parameter not activated on the sensor or sensor still warming up.	Activate parameter or wait until sensor is fully operational.
0x8000 - b15	EP 115 / 8000 Out of range Code 8000 0000 The parameter is out of measurement range	Reading out of measuring range	Measured parameter is outside the defined measuring range.	Check if sensor is in the medium. Perform functional check.

The table below shows all errors regarding clean parameters of the vali::tool software incl. the user message, the reason of the error and notes for trouble shooting. If the error can't be removed although the suggested procedure was executed several times please contact your s::can sales partner.

Parameter Status Error 0xVVVV	Message moni::tool	Reason	Removal
0x0001 - b0	vali::tool reports an error	At least one internal check reports a warning.	Check further status messages.
0x0800 - b11	Maintenance recommended	Parameter check reports a warning.	Check system and sensor, perform functional check.
0x1000 - b12	Marked as not trustable	Parameter check reports a warning.	Do not use this value for calibration.

### 10.3 Device Settings

In case detailed sensor information or configuration settings have to be checked, the following sections will explain how to find these information when operating the sensor with a s::can operation controller.

There is no need to modify any device settings of the chlori::lyser.

#### 10.3.1 Check of Device Settings using con::lyte D-31x

The main menu entry Information of the con::lyte operation software enables you to check internal sensor settings. After selecting the parameter of your interest by pushing the Enter button the display will show the upper limit and the lower limit of the selected parameter.

FCL [mg/l]	
Upper limit:	2.000
Lower limit:	0.000
Span:	0.000
Probe	

When confirming the lowest entry Probe with Enter all internal settings of the chlori::lyser will be displayed. The most important are:

- Internal sensor identifier (M-Version and Model)
- Sensor name (chlori::lyser)
- Serialnumber of the chlori::lyser (S/N)
- Hardware version of the chlori::lyser (H/W-Version)
- Software version of the chlori::lyser (S/W-Version)

#### 10.3.2 Check of Device Settings using con::lyte D-320

Select the entry Manage sensors... in the main menu of the status screen. Select the name chlori::lyser/0/1 in the list of installed sensors, in which the second number (1) indicates the address assigned to the sensor. After confirming the entry Configure... as well as the entry Probesettings in the next view the following information of the sensor will be displayed:

- Internal sensor identifier (M-Version and Model)
- Sensor name (chlori::lyser)
- Serialnumber of the chlori::lyser (S/N)
- Hardware version of the chlori::lyser (H/W-Version)
- Software version of the chlori::lyser (S/W-Version)


P1/FCL	
Sen.:	chlori::lyser
Name:	FCL
Unit:	mg/l
Disp. Format:	3
P. lower:	0,0000
P. upper:	2,0000
Al. lower:	----,---
Al. upper:	----,---

Information of the single measuring parameter can be retrieved via the entry Parameter info... from the main menu of the parameter display. In addition to parameter name (Name), unit of measurement (Unit) the number of decimal places (Disp. Format), also the lower and upper limit of the parameter range (P. lower / P. upper) and the adjusted alarm range (Al. lower / Al. upper) is displayed.

### 10.3.3 Check of Device Settings using moni::tool

Selecting *Service / chlo / Edit chlo* will list up internal settings of the chlори::lyser. Depending on the *Service Level* some or all of the following information will be displayed:


- Interface (COM-port, *Address*) of the sensor
- *Sensor name (Internal)* allocated to the device. Should not be changed by the operator.
- *Sensor name* for the display allocated to the device by the operator at installation.
- Manufacturer name of the sensor (*Vendor*)
- Type of the sensor (*Model*)
- Serial number of the sensor (*Serial Number*)
- Number of internal parameters of the sensor (*Parameter count*)
- Information regarding the purchase (*Purchase date, Warranty expiry date*). Can be entered by the operator at initial start-up.
- Actual hardware and software version of the sensor (*HW Version SW Version*)
- Internal type number of the sensor (*Sensor Model*) and information regarding cleaning and logging (not available for the chlори::lyser)
- Information regarding the installation and last modification of the sensor (date, name and reason)



### 10.3.4 Check of Device Settings using ::pro

In the operating - and service software ::pro the chlори::lyser can be initialised over the menu item *Parameter / Settings*. In case the parameters of interest are not yet displayed by default in the Parameter Settings window, they can be selected as follows:

- Double click on the parameter field that you want to correspond with the parameter to be displayed.
- After double clicking, the window will enlarge and show detailed information belonging to this parameter.
- Select *Non-spectral parameter* in the upper selection bar.
- Select *ammo/chlори/ise::lyser* under *Device*.
- The *COM-Port* is the interface to which the chlори::lyser is connected (e.g. COM-Port of the con::nect).
- Enter the address allocated to the sensor in the RS 485 network (default setting = 9). connection) in the entry *Address*.
- Push the button *Search...* (when an incorrect COM-Port and / or sensor address are selected, the *Search...* function will still find the sensor if only one instrument is connected).





As soon as the sensor has been detected, information will be shown in the grey text field in the lower part of the dialogue window (model and serial number, version, electrodes and measuring range).

Now the internal settings of the chlори::lyser sensor can be checked, but this should be done by s::can Service or after instruction from s::can only.

- Push button Config... in menu Parameter / Settings.
- Push button Search... in register card Com. As soon as the sensor will be detected, the checkbox (status source) in the lower right corner switches to ok (green check mark) and further register cards become visible.
- In register card Device desc the sensor type (abModel), the serial number of the sensor (abSerialNumber), the actual hardware (abHWRelease) and software (abSWRelease) are displayed.
- In register card chlори::lyser config the reference slope (fReferenceSlope) is displayed. This entry should be set to NaN.
- You can finish the configuration menu by pushing the button Exit.

There is no need to modify any device settings of the chlори::lyser.



### 10.4 Return Consignment (RMA)

Return consignments of the s::can monitoring system, or parts of the system, shall be done in a packaging that protects the device (original packaging or protective covering if possible). Before returning a consignment, you have to contact your s::can sales partner or s::can customer support (support@s-can.at). An RMA number will be assigned for each device, independent if the reason of the return consignment is service, repair or demo equipment.

RMA numbers can be requested from the s::can Customer Portal available on the s::can homepage directly. Return consignments without an RMA number will not be accepted. The customer always has to bear the costs for return consignment.

# 11 Accessories

## 11.1 Installation

### 11.1.1 Connection Cable

For operation of chlori::lyser a connection cable is necessary. This is included in the standard order.

Name	Specification	Remark
Part-no.	C-1-010-sensor	
Cable length	1 m	
Assembling	ex works	
Dimensions plug	20 mm	outer diameter
Material	PU	Cable sheathing
Housing environment rating	IP 68	
Interface connection	IP 67, RS 485, 12 VDC	to s::can sensors



### 11.1.2 Extension Cable

The cable of the chlori::lyser can be elongated when necessary with an extension cable (10 m or 20 m length). The extension cable is attached using the sensor cable connector plug.

Name	Specification	Remark
Part-no.	C-210-sensor C-220-sensor	
Cable length	10 m 20 m	C-210-sensor C-220-sensor
Assembling	ex works	
Dimensions plug	20 mm	outer diameter
Material	PU	Cable sheathing
Housing environment rating	IP 68	
Interface connection	IP 67, RS 485, 12 VDC	to s::can sensors



### 11.1.3 Sensor Mounting

For proper and easy submersed installation of the chlori::lyser a separate sensor carrier is available. This part can be fixed to the sensor directly and can be extended by a pipe (to be provided by the customer).




Name	Specification	Remark
Part-no.	F-12-sensor	
Material	PVC, stainless steel	
Dimensions	60 / 91 mm	Diameter / height
Weight	approx. 150 g	
Process connection	DN 50 inside	for extension pipe
Installation / mounting	submersed	




### 11.1.4 Flow Cell Setup Tap Water

For recommended standard installation of chlори:lyser outside the medium, different types of flow cells are available.

Name	Specification	Remark
Part-no.	F-45-sensor F-45-four F-46-four-iscan	for one single sensor for up to four sensors for one i::scan and up to three sensors
Housing material	POM-C	
Dimensions	F-45-sensor: 50 / 96 mm F-45-four: 106 / 103 mm F-46-four-iscan: 106/103mm	Diameter / lenght 177/103 mm with autobrush
Weight	F-45-sensor: ~ 0,25 kg F-45-four: ~ 1.0 kg F-46-four-iscan: ~ 1.0 kg	without autobrush
Process connection	1/4 inch inside	for inlet and outlet
Installation	flow-through	
Mounting	2 mounting holders	
Operating temperature	0 to 50 °C (32 to 122 °F)	
Operating pressure	0 to 6 bar (0 to 87 psi)	



Dimension of flow cell setup in mm (F-45-sensor left side, F-45-four right side)




Dimension of flow cell setup in mm (F-46-four-iscan)

### 11.1.5 Flow Cell Setup Waste Water

For measurement of a waste / raw water sample stream outside the medium with chlори::lyser a separate flow –through installation is available.

Name	Specification	Remark
Part-no.	F-48-sensor	for one single sensor
Housing material	PVC	
Dimensions	177 / 95 / 111 mm	W / H / D
Weight	~ 0,5 kg	
Process connection	1 inch inside (G 1") 40 mm ID	via F-48-process direct connection to G 1"
Installation	flow-through (by-pass)	
Discharge	< 40 l/min	recommended
Operating temperature	0 to 50 °C (32 to 122 °F)	
Operating pressure	0 to 6 bar (0 to 87 psi)	



Dimension of flow cell setup in mm (F-48-sensor)

### 11.1.6 System Panel micro::station / nano::station

For easy attachment of a complete s::can monitoring system (s::can operation terminal, flow cell setup, sensor) different types of separate system panels are available. The process connections of these panels can be ordered in DIN standard (EU) or in National Pipe Standard (US).

Name	Specification	Remark
Part-no.	F-501-eco-xx F-506-panel-xx F-508-panel	Main paneel micro::station Main panel nano::station Waste water panel
Material	PP PE	F-501, F-508 F-506
Dimensions	450 / 750 / 10 mm	F-501-eco
W / H / D	280 / 750 / 10 mm 375 / 690 / 10 mm (per part)	F-506-panel F-508-panel (2 parts)
Weight	~ 4.9 kg	

## 11.2 Spare Parts

### 11.2.1 Serviceset for free Chlorine (E-520-1/2-KIT)

For maintenance of actual version of chlori:lyser for measuring free chlorine, a complete replacement set (electrolyte and membrane cap) is available.

Name	Specification	Remark
Part-no.	E-520-1/2-KIT	for E-520-1 and E-520-2
Scope of delivery	Electrolyte (100ml) Membrane cap (1 pcs.) Polishing paper	
Storage duration		check date of expiry
Storage temperature	10 to 35 °C (50 to 95 °F)	

### 11.2.2 Serviceset for total Chlorine (E-525-1/2-KIT)

For maintenance of actual version of chlori:lyser for measuring total chlorine, a complete replacement set (electrolyte and membrane cap) is available.

Name	Specification	Remark
Part-no.	E-525-1/2-KIT	for E-525-1 and E-525-2
Scope of delivery	Elektrolyte (100ml) Memrankappe (1 pcs.) Schleifpapier	
Storage duration		check date of expiry
Storage temperature	10 to 35 °C (50 to 95 °F)	

### 11.2.3 Electrolyte for free Chlorine (E-507-1/2-EL)


For the actual version (E-520) as well as the previous version (E-507) of the chlori:lyser for measurement of free chlorine, the electrolyte is available in a refill bottle as spare part.

Name	Specification	Remark
Part-no.	E-507-1/2-EL	for E-507-1, E-507-2, E-520-1 and E-520-2
Volume	100 ml	approx. 5 times refilling
Storage duration		check date of expiry
Storage temperature	10 to 35 °C (50 to 95 °F)	

### 11.2.4 Electrolyte for total Chlorine (E-507-3/4-EL)

For the actual version (E-525) as well as the previous version (E-507) of the chlori:lyser for measurement of total chlorine, the electrolyte is available in a refill bottle as spare part.

Name	Specification	Remark
Part-no.	E-507-3/4-EL	for E-507-3, E-507-4, E-525-1 and E-525-2
Volume	100 ml	approx. 5 times refilling
Storage duration		check date of expiry
Storage temperature	10 to 35 °C (50 to 95 °F)	



**s::can** Made in AUSTRIA

scan Messtechnik GmbH  
Brigittgasse 22-24, A-1200

**Electrolyte**  
E-507-1/2-EL, E-520-1/2-EL  
Expiry Date: 2018-05  
Storage Temp.: 10 - 35 °C

Notice:  
For use please refer to the manual

H: 319  
P: 280/264/305+351+338/337+313

**s::can** Made in AUSTRIA

scan Messtechnik GmbH  
Brigittgasse 22-24, A-1200

**Electrolyte**  
E-507-3/4-EL, E-525-1/2-EL  
Expiry Date: 2018-05  
Storage Temp.: 10 - 35 °C

Notice:  
For use please refer to the manual

### 11.2.5 Membrane Cap for free Chlorine (E-507-1/2-SET)

For maintenance of the previous version of chlori::lyser for measuring free chlorine the replaceable membrane cap is available as spare part.

Name	Specification	Remark
Part-no.	E-507-1/2-SET	for E-507-1 and E-507-2
Scope of delivery	Membrane cap (1 pcs.) Spacer (1 pcs.) Polishing paper	
Volume membrane cap	approx. 8 ml Electrolyte	
Storage		only new caps or cleaned caps with max. 24 operating hours can be stored

### 11.2.6 Membrane Cap for total Chlorine (E-507-3/4-SET)


For maintenance of the previous version of the chlori::lyser for measuring free chlorine the replaceable membrane cap is available as spare part.

Name	Specification	Remark
Part-no.	E-507-3/4-SET	for E-507-3 and E-507-4
Scope of delivery	Membrane cap (1 pcs.) Polishing paper	
Volume membrane cap	approx. 8 ml Electrolyte	
Storage		only new caps or cleaned caps with max. 24 operating hours can be stored

## 12 Technical Specifications

Name	Specification E-520-x	Specification E-525-x	Remark
Part-no.	E-520-1 E-520-2	E-525-1 E-525-2	Previous version E-507-x
Measuring parameter	Free chlorine (FCL) Temperature	Total chlorine (TCL) Temperature	see section 5.4
Measuring principle	amperometric 3 electrode sensor, membrane covered		
Compensation	Temperature and pH	Temperature and pH	
Measuring range	E-520-1: 0 - 2 mg/l E-520-2: 0 - 20 mg/l	E-525-1: 0 - 2 mg/l E-525-2: 0 - 20 mg/l	
Resolution	E-520-1: 0.001 mg/l E-520-2: 0.01 mg/l	E-525-1: 0.001 mg/l E-525-2: 0.01 mg/l	
Accuracy	E-520-1: +/- 0.02 at 0.4 +/- 0.02 at 1.6 E-520-2: +/- 0.2 at 4 +/- 0.6 at 16	E-525-1: +/- 0.04 at 0.4 +/- 0.04 at 1.6 E-525-2: +/- 0.2 at 4 +/- 0.6 at 16	in [mg/l] after calibration in drinking water at 25°C and pH 7.2
Reference measurement	DPD-1	DPD-4	for span calibration, zero point factory calibrated
Response time	2 min	2 min	(T <sub>90</sub> )
Running in time (start up)	2 h	2 h	
Installation	in flow cell	in flow cell	
Environment rating	IP 67	IP 67	due to connection plug on sensor
Operating temperature	0 to 45 °C (32 to 113 °F)	0 to 45 °C (32 to 113 °F)	
Operating pressure	0 to 3 bar (0 to 43.5 psi)	0 to 3 bar (0 to 43.5 psi)	no pressure peaks and / or pressure fluctuation
Operating flow	15 to 30 l/h	15 to 30 l/h	low discharge dependency
Operating pH range	4 to 9	4 to 12	
Cross sensitivity	75 % ClO <sub>2</sub> 80 % O <sub>3</sub>  Combined chlorine can increase FCL reading	100 % ClO <sub>2</sub> 130 % O <sub>3</sub>	reducing and oxidizing agents, corrosion inhibitors and stabilisers for water hardness might influence the measurement
Absence of disinfectant	max. 24 h	max. 24 h	biofilm will clog membrane
Power supply	9 to 30 VDC	9 to 30 VDC	permanent supply and output signal galvanically isolated
Power consumption	0.5 W (typ)	0.5 W (typ)	
Dimension	35 / 208 mm 1.38 / 8.19 inch	35 / 208 mm 1.38 / 8.19 inch	Diameter / length (see section 3.3)
Weight	approx. 150 g	approx. 150 g	
Housing material	PVC-U, stainless steel 1.4571		Medium contacted
Interface connection	sys plug (IP 67), RS 485	sys plug (IP 67), RS 485	to s::can operation controller
Sensor cable length	1.0 m	1.0 m	plug connector on top of sensor

Name	Specification E-520-x	Specification E-525-x	Remark
Sensor cable specification	PUR (polyurethane jacket), 22 AWG, 6.3 mm (outside diameter); -30 to 80 °C (-22 to 176 °F)		
Sensor cable assignment	Pin 1: Data - (green cable strand) Pin 2: Data + (pink cable strand) Pin 3: +12 VDC (red cable strand) Pin 4: Ground (black cable strand) Pin 5: not used Pin 6: Shielding (blank cable strand)		
Storage temperature	Sensor: 0 to 45 °C (32 to 113 °F) Electrolyte: 10 to 35 °C (50 to 95 °F)		frost free in original bottle
Storage of sensor	dry, without electrolyte and cleaned membrane cap only		with protective cap to avoid contamination of membrane
Typical lifespan (application)	Membrane: 12 months Electrolyte: 12 months	Membrane: 12 months Electrolyte: 12 months	depending on medium
Typical lifespan (storage)	1 -2 year electrolyte in original bottle, protected from sunlight		check date of expiry on package
Conformity - EMC	EN 61326-1:2013 EN 61326-2-3:2013		General requirements Particular requirements



**scan Messtechnik GmbH**

Brigittagasse 22-24, 1200 Vienna, Austria  
Tel.: +43 (0) 1 219 73 93 - 0  
Fax: +43 (0) 1 219 73 93 - 12  
office@s-can.at  
www.s-can.at

